



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/783,093	02/23/2004	Mordecai Dunst	1954-393	3524
6449	7590	09/21/2007		
ROTHWELL, FIGG, ERNST & MANBECK, P.C. 1425 K STREET, N.W. SUITE 800 WASHINGTON, DC 20005			EXAMINER KASZTEJNA, MATTHEW JOHN	
			ART UNIT 3739	PAPER NUMBER
			NOTIFICATION DATE 09/21/2007	DELIVERY MODE ELECTRONIC

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

PTO-PAT-Email@rfem.com

<b>Office Action Summary</b>	<b>Application No.</b> 10/783,093	<b>Applicant(s)</b> DUNST, MORDECAI	
	<b>Examiner</b> Matthew J. Kasztejna	<b>Art Unit</b> 3739	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

#### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) ☒ Responsive to communication(s) filed on 13 July 2007.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

#### Disposition of Claims

- 4) ☒ Claim(s) 1-6 and 21-49 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-6 and 21-49 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

#### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 23 February 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                                | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                       | 5) <input type="checkbox"/> Notice of Informal Patent Application                       |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## **DETAILED ACTION**

### ***Continued Examination Under 37 CFR 1.114***

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on July 13, 2007 has been entered.

### ***Notice of Amendment***

In response to the amendment filed on July 13, 2007, amended claims 1-3, 26, 28-29, 35 and 37 and new claims 46-49 are acknowledged. The following new grounds of rejection are set forth:

### ***Claim Rejections - 35 USC § 112***

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claim 46 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Claim 46 recites the limitation "the proximal" in line 10 of the claim. There is insufficient antecedent basis for this limitation in the claim.

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-2, 4, 6, 21-28, 30-34 are rejected under 35 U.S.C. 102(b) as being anticipated by U.S. Patent No. 2,911,968 to Schueler et al.

**In regards to claims 1 and 27**, Schueler et al. disclose a blade assembly for a laryngoscope including a handle 1, the blade assembly comprising; base for coupling to the handle; an elongated blade 3 secured to the base, the elongated blade having a proximal end, a distal end, an upper surfaces and a lower surface, the lower surface of the elongated blade being secured to the base at the proximal end; a first guide tube 7 (or 48) affixed to the elongated blade and having a proximal end and a distal end; and a second guide tube 52 affixed to the elongated blade adjacent the first guide tube and having a proximal end and a distal end, the first and second guide tubes being constructed and arranged to direct tubes extended therethrough into the oral cavity of a patient (see Figs. 1-3).

**In regards to claims 2 and 28**, Schueler et al. discloses a blade assembly for a laryngoscope, wherein the first guide tube 7 is substantially straight along its entire length and the second guide tube is curved 52 (see Fig. 1).

**In regards to claim 4**, Schueler et al. discloses assembly for a laryngoscope, further comprising a light attached to the elongated blade (see Col. 2, Lines 40-65).

**In regards to claims 6 and 34**, Schueler et al. discloses a blade assembly for a laryngoscope, wherein the first and second guide tubes are oriented at different angles of attack 53, 50 with respect to the elongated blade so that when the elongated blade is placed atop the tongue of a patient to effect laryngeal suspension, the first end second guide tubes direct respective tubes extended therethrough into different regions of the patient's oral cavity (see Col. 4, Lines 21-35).

**In regards to claims 21 and 30**, Schueler et al. discloses a blade assembly for a laryngoscope, wherein the proximal end of the first guide tube and the proximal end of the second guide tube are both positioned at the proximal end of the elongated blade (see Figs. 1-2).

**In regards to claims 22-24 and 31-32**, Schueler et al. discloses a blade assembly for a laryngoscope, wherein the distal end of the guide tube and the distal end of the second guide tube are positioned at about the midpoint of the elongated blade (See Fig. 1).

**In regards to claims 25-26 and 33**, Schueler et al. discloses a blade assembly for a laryngoscope, further comprising a first aspiration tube inserted into the first guide tube and a second aspiration tube inserted into the second guide tube (see Col. 4, Lines 21-35).

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the

Art Unit: 3739

invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 3, 5, 29 are rejected under 35 U.S.C. 103(a) as being unpatentable over U.S. Patent No. 2,911,968 to Schueler et al. in view of U.S. Patent No. 6,248,061 to Cook, Jr.

**In regards to claims 3, 5 and 29,** Schueler et al. disclose a blade assembly for a laryngoscope including a handle 1, the blade assembly comprising; base for coupling to the handle; an elongated blade 3 secured to the base, the elongated blade having a proximal end, a distal end, an upper surfaces and a lower surface, the lower surface of the elongated blade being secured to the base at the proximal end; a first guide tube 7 (or 48) affixed to the elongated blade and having a proximal end and a distal end; and a second guide tube 52 affixed to the elongated blade adjacent the first guide tube and having a proximal end and a distal end, the first and second guide tubes being constructed and arranged to direct tubes extended therethrough into the oral cavity of a patient (see Figs. 1-3) but is silent with respect to the elongated blade being curved. Cook, Jr. teach of an analogous laryngoscope blade 60 having a curved blade portion 62, a proximal end 66 and a distal end 64 and a base portion 74 to connect to a handle 26 (see Figure 6). Figures 6 and 7 show that the blade portion 62 has an upper surface and a lower surface and that the lower surface is secured to the base portion 74 at the proximal end 66. It would have been obvious to one skilled in the art at the time the invention was made to curve the elongate blade of Schueler et al. to aid in the placement of the laryngoscope within the airway of a patient as taught by Cook, Jr. and is well known in the art.

Art Unit: 3739

Claims 35-45 are rejected under 35 U.S.C. 103(a) as being unpatentable over U.S. Patent No. 6,248,061 to Cook, Jr. in view of U.S. Patent No. 2,911,968 to Schueler et al.

In regard to claims 35-45, Cook, Jr. teach a laryngoscope blade 60 having a curved blade portion 62, a proximal end 66 and a distal end 64 and a base portion 74 to connect to a handle 26 (see Figure 6). Figures 6 and 7 show that the blade portion 62 has an upper surface and a lower surface and that the lower surface is secured to the base portion 74 at the proximal end 66. A suction tube 81 is secured to the upper surface of blade portion 62 and is constructed and arranged to direct suction extended therethrough into the oral cavity of a patient (see Figures 6 and 7). Furthermore, Cook, Jr. teaches a light source 72 attached to blade 60 (see Figures 6 and 7). In regard to claim 5, blade 60 is provided with a side wall 68 and a fin 70 that extend in a direction transverse to the upper surface of blade 60 along a marginal edge segment (see Figures 6 and 7). However, Cook, Jr. is silent as to a second suction tube. Schueler et al. teach of an analogous laryngoscope having first and second guide tubes 48 and 51 (see Col. 4, Lines 21-35 and Fig. 1). It would have been obvious to one skilled in the art at the time the invention was made to have a second guide tube on secured to the upper blade of curve the elongate blade of Cook, Jr. in order to direct suction into different regions of the patient's oral cavity as taught by Schueler et al.

***Allowable Subject Matter***

Claim 46 would be allowable if rewritten or amended to overcome the rejection(s) under 35 U.S.C. 112, 2nd paragraph, set forth in this Office action.

Art Unit: 3739

Claims 47-49 would be allowable if rewritten to overcome the rejection(s) under 35 U.S.C. 112, 2nd paragraph, set forth in this Office action and to include all of the limitations of the base claim and any intervening claims.

***Response to Arguments***

Applicant's arguments with respect to claims 1-6 and 21-49 have been considered but are moot in view of the new ground(s) of rejection.

***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Matthew J. Kasztejna whose telephone number is (571) 272-6086. The examiner can normally be reached on Mon-Fri, 8:30-6:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C.M. Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.




Art Unit: 3739

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MJK *me*

9/13/07



LINDA C. M. DVORAK  
SUPERVISORY PATENT EXAMINER  
GROUP 3700